

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062018\
 Data File : VN049457.D
 Acq On : 21 Jun 2018 00:22
 Operator : MD\SY
 Sample : J3584-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3

Quant Time: Jun 21 03:24:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1088846	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1731587	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1405154	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	424228	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	746676	51.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.40%	
35) Dibromofluoromethane	7.59	113	644308	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
50) Toluene-d8	10.09	98	2345384	43.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.94%	
62) 4-Bromofluorobenzene	12.40	95	598463	33.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.86%	
Target Compounds						
16) Acetone	3.82	43	36460	10.40	ug/l	95
18) Methyl Acetate	4.33	43	92289	8.63	ug/l	93
20) Methylene Chloride	4.55	84	188155	12.85	ug/l	93
84) 1,2,4-Trimethylbenzene	13.04	105	52023	1.79	ug/l	99
95) Naphthalene	15.14	128	200586	18.35	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062018\
Data File : VN049457.D
Acq On : 21 Jun 2018 00:22
Operator : MD\SY
Sample : J3584-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: Jun 21 03:24:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

