

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062118\
 Data File : VN049467.D
 Acq On : 21 Jun 2018 14:35
 Operator : MD\SY
 Sample : PB110441ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110441ZHE#04

Quant Time: Jun 22 04:09:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1153197	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1906190	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1550286	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	415919	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	807680	52.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.62%	
35) Dibromofluoromethane	7.59	113	720429	46.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.42%	
50) Toluene-d8	10.09	98	2550832	43.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.88%	
62) 4-Bromofluorobenzene	12.40	95	628875	32.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	64.78%	
Target Compounds						
16) Acetone	3.82	43	30857	8.31	ug/l	# 87
18) Methyl Acetate	4.33	43	86891	7.67	ug/l	97
20) Methylene Chloride	4.55	84	47413	3.06	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062118\
Data File : VN049467.D
Acq On : 21 Jun 2018 14:35
Operator : MD\SY
Sample : PB110441ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110441ZHE#04

Quant Time: Jun 22 04:09:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

