

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062118\
 Data File : VN049476.D
 Acq On : 21 Jun 2018 18:45
 Operator : MD\SY
 Sample : PB110441ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110441ZHE#07

Quant Time: Jun 22 04:43:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1177752	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1857286	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1558074	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	422359	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	798621	51.12	ug/l	0.00
Spiked Amount			Recovery	=	102.24%	
35) Dibromofluoromethane	7.59	113	686851	45.71	ug/l	0.00
Spiked Amount			Recovery	=	91.42%	
50) Toluene-d8	10.09	98	2552503	44.61	ug/l	0.00
Spiked Amount			Recovery	=	89.22%	
62) 4-Bromofluorobenzene	12.41	95	635235	33.58	ug/l	0.00
Spiked Amount			Recovery	=	67.16%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	31595	8.33	ug/l	92
18) Methyl Acetate	4.33	43	89482	7.74	ug/l	95
20) Methylene Chloride	4.56	84	40341	2.55	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062118\
Data File : VN049476.D
Acq On : 21 Jun 2018 18:45
Operator : MD\SY
Sample : PB110441ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110441ZHE#07

Quant Time: Jun 22 04:43:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

