

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062118\
 Data File : VN049483.D
 Acq On : 21 Jun 2018 21:45
 Operator : MD\SY
 Sample : J3622-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP-A

Quant Time: Jun 22 05:13:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1280761	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1992424	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1667113	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	524111	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	852504	50.19	ug/l	0.00
Spiked Amount			Recovery	=	100.38%	
35) Dibromofluoromethane	7.59	113	736685	45.70	ug/l	0.00
Spiked Amount			Recovery	=	91.40%	
50) Toluene-d8	10.09	98	2761235	44.99	ug/l	0.00
Spiked Amount			Recovery	=	89.98%	
62) 4-Bromofluorobenzene	12.40	95	734854	36.21	ug/l	0.00
Spiked Amount			Recovery	=	72.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	28758	6.97	ug/l	98
18) Methyl Acetate	4.34	43	86834	6.90	ug/l	96
20) Methylene Chloride	4.55	84	1229403	71.37	ug/l	92
95) Naphthalene	15.14	128	182155	14.19	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062118\
Data File : VN049483.D
Acq On : 21 Jun 2018 21:45
Operator : MD\SY
Sample : J3622-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP-A

Quant Time: Jun 22 05:13:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

