


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN062118\
 Data File : VN049460.D
 Acq On : 21 Jun 2018 10:30
 Operator : MD\SY
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Title : SW846 8260
 Last Update : Sat Jun 16 01:02:01 2018



AutoFind: Scans 3363, 3364, 3365; Background Corrected with Scan 3348

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.7	19618	PASS
75	95	30	60	49.6	54952	PASS
95	95	100	100	100.0	110882	PASS
96	95	5	9	6.2	6828	PASS
173	174	0.00	2	1.3	1360	PASS
174	95	50	100	91.8	101816	PASS
175	174	5	9	7.5	7596	PASS
176	174	95	101	95.1	96800	PASS
177	176	5	9	6.2	6028	PASS