

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062824\
 Data File : VN082742.D
 Acq On : 28 Jun 2024 11:02
 Operator : JC\MD
 Sample : VN0628WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0628WBL01

Quant Time: Jun 28 23:01:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 06:02:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	75382	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	143044	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	137987	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	54272	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	64434	55.427	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.860%	
35) Dibromofluoromethane	8.165	113	52212	55.439	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.880%	
50) Toluene-d8	10.570	98	185039	52.537	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.080%	
62) 4-Bromofluorobenzene	12.853	95	57342	47.408	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 94.820%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062824\
Data File : VN082742.D
Acq On : 28 Jun 2024 11:02
Operator : JC\MD
Sample : VN0628WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0628WBL01

Quant Time: Jun 28 23:01:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 06:02:21 2024
Response via : Initial Calibration

