

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062824\
 Data File : VN082747.D
 Acq On : 28 Jun 2024 14:58
 Operator : JC\MD
 Sample : PB161828TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161828TB

Quant Time: Jun 28 23:04:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 06:02:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	69239	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	139065	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	150972	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	61458	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	61199	57.315	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 114.620%			
35) Dibromofluoromethane	8.165	113	48572	53.049	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.100%			
50) Toluene-d8	10.571	98	192863	56.325	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 112.640%			
62) 4-Bromofluorobenzene	12.853	95	62755	53.367	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 106.740%			
Target Compounds						Qvalue
16) Acetone	4.430	43	23859	58.976	ug/l	91
20) Methylene Chloride	5.277	84	4498	4.394	ug/l	85
43) Isopropyl Acetate	8.688	43	67593	27.406	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062824\
Data File : VN082747.D
Acq On : 28 Jun 2024 14:58
Operator : JC\MD
Sample : PB161828TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161828TB

Quant Time: Jun 28 23:04:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 06:02:21 2024
Response via : Initial Calibration

