

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062824\
 Data File : VN082749.D
 Acq On : 28 Jun 2024 16:36
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: Jun 28 23:14:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 06:02:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	69594	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	135325	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	134250	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	54153	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	59402	55.348	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.700%
35) Dibromofluoromethane	8.165	113	48091	53.976	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.960%
50) Toluene-d8	10.571	98	175674	52.723	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.440%
62) 4-Bromofluorobenzene	12.853	95	56178	49.094	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	98.180%
Target Compounds						
25) 2-Butanone	7.489	43	21275	34.246	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062824\
Data File : VN082749.D
Acq On : 28 Jun 2024 16:36
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: Jun 28 23:14:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 06:02:21 2024
Response via : Initial Calibration

