

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062824\
 Data File : VN082750.D
 Acq On : 28 Jun 2024 17:00
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: Jun 28 23:14:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 06:02:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	67947	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	129251	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	128957	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	52419	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	57801	55.162	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.320%
35) Dibromofluoromethane	8.165	113	47962	56.361	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.720%
50) Toluene-d8	10.571	98	170258	53.498	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.000%
62) 4-Bromofluorobenzene	12.853	95	52879	48.383	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	96.760%
Target Compounds						
25) 2-Butanone	7.494	43	3957	6.524	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062824\
Data File : VN082750.D
Acq On : 28 Jun 2024 17:00
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: Jun 28 23:14:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 06:02:21 2024
Response via : Initial Calibration

