

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071825\
 Data File : VN087363.D
 Acq On : 18 Jul 2025 13:59
 Operator : JC\MD
 Sample : Q2630-02DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 002 35th Ave (JUNE)DL

Quant Time: Jul 18 14:33:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N062525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jun 25 10:49:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.800	128	45506	30.000	ug/l	-0.02
28) 1,4-Difluorobenzene	9.082	114	211698	30.000	ug/l	-0.02
57) Chlorobenzene-d5	11.847	117	193269	30.000	ug/l	-0.02
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.565	65	100699	29.370	ug/l	-0.02
Spiked Amount	30.000	Range	91 - 110	Recovery	=	97.900%
60) 4-Bromofluorobenzene	12.829	95	83753	28.083	ug/l	-0.02
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.600%
63) Toluene-d8	10.547	98	257514	29.675	ug/l	-0.02
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.933%
Target Compounds						
					Qvalue	
15) Acetone	4.436	58	25861	43.958	ug/l	94
30) 2-Butanone	7.477	43	26473	11.894	ug/l #	81
62) Toluene	10.612	91	896943	82.171	ug/l	99
82) p-Isopropyltoluene	13.706	119	11453	1.272	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071825\
Data File : VN087363.D
Acq On : 18 Jul 2025 13:59
Operator : JC\MD
Sample : Q2630-02DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
002 35th Ave (JUNE)DL

Quant Time: Jul 18 14:33:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N062525W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jun 25 10:49:56 2025
Response via : Initial Calibration

