

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072821\
 Data File : VN067953.D
 Acq On : 28 Jul 2021 15:13
 Operator : JC/MD
 Sample : M3195-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SUMPWATER-TRE-1232

Quant Time: Jul 29 02:51:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	301500	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	566110	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	538560	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	200552	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	231557	50.818	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 101.640%			
35) Dibromofluoromethane	8.024	113	175418	51.287	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 102.580%			
50) Toluene-d8	10.443	98	710502	49.628	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 99.260%			
62) 4-Bromofluorobenzene	12.733	95	236349	46.733	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 93.460%			
Target Compounds						
16) Acetone	4.296	43	10305	7.187	ug/l	99
29) Tetrahydrofuran	7.694	42	61975	39.584	ug/l	93
69) o-Xylene	12.283	106	11938	1.594	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN072821\
Data File : VN067953.D
Acq On : 28 Jul 2021 15:13
Operator : JC/MD
Sample : M3195-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SUMPWATER-TRE-1232

Quant Time: Jul 29 02:51:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

