

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
 Data File : VN083260.D
 Acq On : 13 Aug 2024 16:07
 Operator : JC\MD
 Sample : P3390-07 2.5PPB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Aug 14 03:01:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 14 02:57:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	164141	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	287138	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	245212	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	109492	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.577	65	129179	55.291	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.580%	
35) Dibromofluoromethane	8.165	113	91753	51.194	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.380%	
50) Toluene-d8	10.571	98	338557	50.640	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.280%	
62) 4-Bromofluorobenzene	12.847	95	125867	48.291	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 96.580%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.124	85	4787	2.572	ug/l	96
3) Chloromethane	2.365	50	5745	3.015	ug/l	90
4) Vinyl Chloride	2.512	62	5550	2.854	ug/l #	88
5) Bromomethane	2.953	94	4066	3.370	ug/l	97
6) Chloroethane	3.112	64	3566	2.931	ug/l	98
7) Trichlorofluoromethane	3.489	101	8907	2.773	ug/l	96
8) Diethyl Ether	3.959	74	3005	2.514	ug/l	85
9) 1,1,2-Trichlorotrifluo...	4.371	101	5247	2.963	ug/l #	90
10) Methyl Iodide	4.589	142	5194	2.229	ug/l	92
11) Tert butyl alcohol	5.530	59	5683	11.705	ug/l #	90
12) 1,1-Dichloroethene	4.336	96	4508	2.478	ug/l	87
13) Acrolein	4.183	56	3650	11.534	ug/l	98
14) Allyl chloride	5.024	41	8265	2.404	ug/l	91
15) Acrylonitrile	5.724	53	12484	12.502	ug/l	97
16) Acetone	4.442	43	12059	13.191	ug/l	95
17) Carbon Disulfide	4.712	76	12618	2.370	ug/l	100
18) Methyl Acetate	5.042	43	7194	2.641	ug/l	91
19) Methyl tert-butyl Ether	5.789	73	16912	2.575	ug/l	97
20) Methylene Chloride	5.277	84	6227	2.960	ug/l #	86
21) trans-1,2-Dichloroethene	5.794	96	5168	2.748	ug/l	83
22) Diisopropyl ether	6.677	45	16505	2.554	ug/l	97
23) Vinyl Acetate	6.600	43	88998	13.434	ug/l	95
24) 1,1-Dichloroethane	6.565	63	9717	2.758	ug/l #	93
25) 2-Butanone	7.488	43	17757	12.650	ug/l	97
26) 2,2-Dichloropropane	7.500	77	7768	2.374	ug/l	93
27) cis-1,2-Dichloroethene	7.488	96	6090	2.684	ug/l	94
28) Bromochloromethane	7.812	49	3847	2.672	ug/l #	87
29) Tetrahydrofuran	7.853	42	11527	12.703	ug/l #	88
30) Chloroform	7.965	83	9380	2.563	ug/l	96
31) Cyclohexane	8.259	56	12370	3.573	ug/l	93
32) 1,1,1-Trichloroethane	8.171	97	9151	2.641	ug/l #	52
36) 1,1-Dichloropropene	8.377	75	7031	2.593	ug/l	96
37) Ethyl Acetate	7.565	43	7769	2.559	ug/l #	92
38) Carbon Tetrachloride	8.359	117	8074	2.644	ug/l	93
39) Methylcyclohexane	9.600	83	8010	2.405	ug/l	98
40) Benzene	8.606	78	20880	2.585	ug/l	93

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
 Data File : VN083260.D
 Acq On : 13 Aug 2024 16:07
 Operator : JC\MD
 Sample : P3390-07 2.5PPB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Aug 14 03:01:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 14 02:57:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.783	41	4434	2.569	ug/l	97
42) 1,2-Dichloroethane	8.671	62	8069	2.743	ug/l	97
43) Isopropyl Acetate	8.694	43	17110	2.460	ug/l	93
44) Trichloroethene	9.359	130	5254	2.733	ug/l	95
45) 1,2-Dichloropropane	9.629	63	5360	2.796	ug/l #	86
46) Dibromomethane	9.706	93	3583	2.611	ug/l	96
47) Bromodichloromethane	9.882	83	8189	2.658	ug/l #	89
48) Methyl methacrylate	9.682	41	5712	2.276	ug/l	91
49) 1,4-Dioxane	9.706	88	2189	48.347	ug/l	95
51) 4-Methyl-2-Pentanone	10.447	43	36307	12.648	ug/l	96
52) Toluene	10.629	92	13175	2.582	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	7457	2.356	ug/l	95
54) cis-1,3-Dichloropropene	10.312	75	7645	2.271	ug/l #	91
55) 1,1,2-Trichloroethane	11.018	97	4906	2.683	ug/l	90
56) Ethyl methacrylate	10.876	69	7886	2.287	ug/l	93
57) 1,3-Dichloropropane	11.165	76	8343	2.559	ug/l	96
58) 2-Chloroethyl Vinyl ether	10.165	63	18155	12.458	ug/l	98
59) 2-Hexanone	11.194	43	26450	11.910	ug/l	92
60) Dibromochloromethane	11.359	129	5422	2.452	ug/l	95
61) 1,2-Dibromoethane	11.471	107	4819	2.508	ug/l	96
64) Tetrachloroethene	11.106	164	4218	2.597	ug/l	94
65) Chlorobenzene	11.894	112	14468	2.670	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.959	131	4653	2.435	ug/l	91
67) Ethyl Benzene	11.965	91	25209	2.536	ug/l	95
68) m/p-Xylenes	12.070	106	18254	4.902	ug/l	96
69) o-Xylene	12.400	106	8632	2.350	ug/l	94
70) Styrene	12.412	104	14755	2.392	ug/l	99
71) Bromoform	12.582	173	3676	2.539	ug/l #	97
73) Isopropylbenzene	12.700	105	22585	2.466	ug/l	99
74) N-amyl acetate	12.494	43	10079	2.250	ug/l	95
75) 1,1,2,2-Tetrachloroethane	12.941	83	6834	2.638	ug/l	96
76) 1,2,3-Trichloropropane	12.994	75	9018	3.753	ug/l	51
77) Bromobenzene	12.976	156	5448	2.678	ug/l	93
78) n-propylbenzene	13.035	91	26554	2.518	ug/l	99
79) 2-Chlorotoluene	13.123	91	17153	2.566	ug/l	100
80) 1,3,5-Trimethylbenzene	13.170	105	18888	2.463	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	2334	2.113	ug/l	92
82) 4-Chlorotoluene	13.223	91	16553	2.468	ug/l	99
83) tert-Butylbenzene	13.435	119	16371	2.411	ug/l	95
84) 1,2,4-Trimethylbenzene	13.482	105	18418	2.384	ug/l	99
85) sec-Butylbenzene	13.617	105	22194	2.395	ug/l	100
86) p-Isopropyltoluene	13.729	119	17129	2.239	ug/l	95
87) 1,3-Dichlorobenzene	13.735	146	9844	2.572	ug/l	97
88) 1,4-Dichlorobenzene	13.735	146	9844	2.551	ug/l	95
89) n-Butylbenzene	14.059	91	14656	2.211	ug/l	99
90) Hexachloroethane	14.329	117	3712	2.512	ug/l	76
91) 1,2-Dichlorobenzene	14.112	146	9107	2.458	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.723	75	1315	2.092	ug/l	95
93) 1,2,4-Trichlorobenzene	15.394	180	4513	2.175	ug/l	97
94) Hexachlorobutadiene	15.494	225	1985	2.151	ug/l	99
95) Naphthalene	15.647	128	15084	2.052	ug/l	98
96) 1,2,3-Trichlorobenzene	15.841	180	4405	2.145	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
Data File : VN083260.D
Acq On : 13 Aug 2024 16:07
Operator : JC\MD
Sample : P3390-07 2.5PPB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Aug 14 03:01:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 14 02:57:40 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:168 Resp: 164141
Ion Ratio Lower Upper
168 100
99 66.5 48.2 72.4



#2
Dichlorodifluoromethane
Concen: 2.572 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Tgt Ion: 85 Resp: 4787
Ion Ratio Lower Upper
85 100
87 33.6 15.7 47.0









#3
Chloromethane
Concen: 3.015 ug/l
RT: 2.365 min Scan# 70
Delta R.T. 0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 50 Resp: 5745
Ion Ratio Lower Upper
50 100
52 36.9 25.0 37.4



#4
Vinyl Chloride
Concen: 2.854 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07


Tgt Ion: 62 Resp: 5550
Ion Ratio Lower Upper
62 100
64 24.8 25.0 37.6#








#5
Bromomethane
Concen: 3.370 ug/l
RT: 2.953 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 94 Resp: 4066
Ion Ratio Lower Upper
94 100
96 94.4 78.0 117.0



#6
Chloroethane
Concen: 2.931 ug/l
RT: 3.112 min Scan# 197
Delta R.T. 0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Tgt Ion: 64 Resp: 3566
Ion Ratio Lower Upper
64 100
66 32.1 26.6 40.0









#7
Trichlorofluoromethane
Concen: 2.773 ug/l
RT: 3.489 min Scan# 210
Delta R.T. -0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:101 Resp: 8907
Ion Ratio Lower Upper
101 100
103 62.4 52.7 79.1



#8
Diethyl Ether
Concen: 2.514 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07


Tgt Ion: 74 Resp: 3005
Ion Ratio Lower Upper
74 100
45 94.6 55.5 166.3








#9
1,1,2-Trichlorotrifluoroethane
Concen: 2.963 ug/l
RT: 4.371 min Scan# 410
Delta R.T. 0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:101 Resp: 5247
Ion Ratio Lower Upper
101 100
85 38.2 38.3 57.5#
151 68.3 59.3 88.9



#10
Methyl Iodide
Concen: 2.229 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Tgt Ion:142 Resp: 5194
Ion Ratio Lower Upper
142 100
127 41.4 37.5 56.3
141 14.0 13.1 19.7









#11
Tert butyl alcohol
Concen: 11.705 ug/l
RT: 5.530 min Scan# 608
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 59 Resp: 5683
Ion Ratio Lower Upper
59 100
57 7.2 8.9 13.3#



#12
1,1-Dichloroethene
Concen: 2.478 ug/l
RT: 4.336 min Scan# 405
Delta R.T. -0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Tgt Ion: 96 Resp: 4508
Ion Ratio Lower Upper
96 100
61 166.4 149.7 224.5
98 56.5 50.1 75.1





#13

Acrolein

Concen: 11.534 ug/l

RT: 4.183 min Scan# 31

Delta R.T. 0.000 min

Lab File: VN083260.D

Acq: 13 Aug 2024 16:07

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 56 Resp: 3650

Ion Ratio Lower Upper

56 100

55 71.9 56.4 84.6



#14

Allyl chloride

Concen: 2.404 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN083260.D

Acq: 13 Aug 2024 16:07




Tgt Ion: 41 Resp: 8265

Ion Ratio Lower Upper

41 100

39 73.9 67.8 101.6


76 32.9 25.8 38.8








#15
Acrylonitrile
Concen: 12.502 ug/l
RT: 5.724 min Scan# 641
Delta R.T. 0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion: 53 Resp: 12484
Ion Ratio Lower Upper
53 100
52 80.7 66.6 100.0
51 36.4 30.1 45.1




#16
Acetone
Concen: 13.191 ug/l
RT: 4.442 min Scan# 423
Delta R.T. 0.012 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07



Tgt Ion: 43 Resp: 12059
Ion Ratio Lower Upper
43 100
58 29.9 21.8 32.6









#17
Carbon Disulfide
Concen: 2.370 ug/l
RT: 4.712 min Scan# 41
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 76 Resp: 12618
Ion Ratio Lower Upper
76 100
78 9.1 7.2 10.8



#18
Methyl Acetate
Concen: 2.641 ug/l
RT: 5.042 min Scan# 525
Delta R.T. 0.018 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07


Tgt Ion: 43 Resp: 7194
Ion Ratio Lower Upper
43 100
74 24.4 16.3 24.5








#19
Methyl tert-butyl Ether
Concen: 2.575 ug/l
RT: 5.789 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 73 Resp: 16912
Ion Ratio Lower Upper
73 100
57 21.0 17.9 26.9



#20
Methylene Chloride
Concen: 2.960 ug/l
RT: 5.277 min Scan# 565
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Tgt Ion: 84 Resp: 6227
Ion Ratio Lower Upper
84 100
49 125.1 119.6 179.4
51 31.4 34.8 52.2#
86 66.6 52.9 79.3









#21
trans-1,2-Dichloroethene
Concen: 2.748 ug/l
RT: 5.794 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 96 Resp: 5168
Ion Ratio Lower Upper
96 100
61 129.6 123.4 185.2
98 52.3 49.3 73.9



#22
Diisopropyl ether
Concen: 2.554 ug/l
RT: 6.677 min Scan# 803
Delta R.T. 0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Tgt Ion: 45 Resp: 16505
Ion Ratio Lower Upper
45 100
43 57.9 44.0 66.0
87 23.2 19.7 29.5
59 10.8 8.5 12.7









#23
 Vinyl Acetate
 Concen: 13.434 ug/l
 RT: 6.600 min Scan# 791
 Delta R.T. -0.006 min
 Lab File: VN083260.D
 Acq: 13 Aug 2024 16:07




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 43 Resp: 88998
 Ion Ratio Lower Upper
 43 100
 86 10.9 7.3 10.9



#24
 1,1-Dichloroethane
 Concen: 2.758 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. -0.006 min
 Lab File: VN083260.D
 Acq: 13 Aug 2024 16:07


Tgt Ion: 63 Resp: 9717
 Ion Ratio Lower Upper
 63 100
 98 7.7 3.3 9.9
 100 0.0 2.0 6.0#








#25
2-Butanone
Concen: 12.650 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 43 Resp: 17757
Ion Ratio Lower Upper
43 100
72 24.1 18.2 27.2



#26
2,2-Dichloropropane
Concen: 2.374 ug/l
RT: 7.500 min Scan# 943
Delta R.T. 0.012 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07


Tgt Ion: 77 Resp: 7768
Ion Ratio Lower Upper
77 100
97 17.6 10.3 30.9








#27
 cis-1,2-Dichloroethene
 Concen: 2.684 ug/l
 RT: 7.488 min Scan# 941
 Delta R.T. 0.000 min
 Lab File: VN083260.D
 Acq: 13 Aug 2024 16:07

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion:	96	Resp:	6090
Ion	Ratio	Lower	Upper
96	100		
61	155.1	0.0	329.6
98	61.3	0.0	128.6



#28
 Bromochloromethane
 Concen: 2.672 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. -0.000 min
 Lab File: VN083260.D
 Acq: 13 Aug 2024 16:07

Tgt Ion:	49	Resp:	3847
Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	3.8
130	73.1	50.5	75.7





#29

Tetrahydrofuran

Concen: 12.703 ug/l

RT: 7.853 min Scan# 1003

Delta R.T. 0.018 min

Lab File: VN083260.D





Acq: 13 Aug 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion:	42	Resp:	11527
Ion Ratio	Lower	Upper	
42	100		
72	43.7	30.0	45.0
71	43.8	28.3	42.5#



#30

Chloroform

Concen: 2.563 ug/l




RT: 7.965 min Scan# 1022


Delta R.T. -0.000 min

Lab File: VN083260.D

Acq: 13 Aug 2024 16:07


Tgt Ion:	83	Resp:	9380
Ion Ratio	Lower	Upper	
83	100		
85	66.5	50.9	76.3








#31
Cyclohexane
Concen: 3.573 ug/l
RT: 8.259 min Scan# 1072
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 56 Resp: 12370
Ion Ratio Lower Upper
56 100
69 31.8 24.5 36.7
84 74.3 66.0 99.0



#32
1,1,1-Trichloroethane
Concen: 2.641 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Tgt Ion: 97 Resp: 9151
Ion Ratio Lower Upper
97 100
99 0.0 52.0 78.0#
61 48.0 42.1 63.1





#33

1,2-Dichloroethane-d4

Concen: 55.291 ug/l

RT: 8.577 min Scan# 1127

Delta R.T. -0.006 min

Lab File: VN083260.D

Acq: 13 Aug 2024 16:07

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 65 Resp: 129179

Ion Ratio Lower Upper

65 100

67 50.9 0.0 104.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN083260.D

Acq: 13 Aug 2024 16:07




Tgt Ion: 114 Resp: 287138


Ion Ratio Lower Upper

114 100

63 22.5 0.0 44.6


88 16.9 0.0 31.4








#35
Dibromofluoromethane
Concen: 51.194 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:113 Resp: 91753
Ion Ratio Lower Upper
113 100
111 103.3 82.4 123.6
192 17.4 14.9 22.3



#36
1,1-Dichloropropene
Concen: 2.593 ug/l
RT: 8.377 min Scan# 1092
Delta R.T. 0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Tgt Ion: 75 Resp: 7031
Ion Ratio Lower Upper
75 100
110 34.6 16.3 48.9
77 28.2 24.6 37.0





#37

Ethyl Acetate

Concen: 2.559 ug/l

RT: 7.565 min Scan# 91

Delta R.T. 0.006 min


Lab File: VN083260.D

Acq: 13 Aug 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion: 43 Resp: 7769

Ion Ratio Lower Upper

43 100

61 8.3 9.3 13.9#

70 11.6 7.0 10.6#



#38

Carbon Tetrachloride

Concen: 2.644 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. 0.000 min

Lab File: VN083260.D

Acq: 13 Aug 2024 16:07




Tgt Ion: 117 Resp: 8074


Ion Ratio Lower Upper

117 100

119 85.5 74.9 112.3

121 32.2 24.3 36.5





#39

Methylcyclohexane

Concen: 2.405 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. -0.000 min


Lab File: VN083260.D

Acq: 13 Aug 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion: 83 Resp: 8010

Ion Ratio Lower Upper

83 100

55 83.6 66.4 99.6

98 47.7 40.7 61.1



#40

Benzene

Concen: 2.585 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN083260.D




Acq: 13 Aug 2024 16:07


Tgt Ion: 78 Resp: 20880

Ion Ratio Lower Upper

78 100

77 27.4 19.0 28.4









#41
Methacrylonitrile
Concen: 2.569 ug/l
RT: 7.783 min Scan# 990
Delta R.T. 0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:	41	Resp:	4434
Ion	Ratio	Lower	Upper
41	100		
39	56.7	45.5	68.3
67	59.3	50.5	75.7
52	27.3	24.0	36.0



#42
1,2-Dichloroethane
Concen: 2.743 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07


Tgt Ion:	62	Resp:	8069
Ion	Ratio	Lower	Upper
62	100		
98	6.9	0.0	15.8








#43
Isopropyl Acetate
Concen: 2.460 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. 0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 43 Resp: 17110
Ion Ratio Lower Upper
43 100
61 18.4 17.8 26.6
87 8.5 8.2 12.2



#44
Trichloroethene
Concen: 2.733 ug/l
RT: 9.359 min Scan# 1259
Delta R.T. 0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Tgt Ion: 130 Resp: 5254
Ion Ratio Lower Upper
130 100
95 104.3 0.0 197.8





#45

1,2-Dichloropropane

Concen: 2.796 ug/l

RT: 9.629 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN083260.D

Acq: 13 Aug 2024 16:07

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 63 Resp: 5360

Ion Ratio Lower Upper

63 100

65 22.6 24.4 36.6#



#46

Dibromomethane

Concen: 2.611 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.006 min

Lab File: VN083260.D

Acq: 13 Aug 2024 16:07




Tgt Ion: 93 Resp: 3583


Ion Ratio Lower Upper

93 100

95 83.2 65.8 98.6

174 95.0 71.7 107.5









#47
Bromodichloromethane
Concen: 2.658 ug/l
RT: 9.882 min Scan# 1349
Delta R.T. -0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 8189
Ion Ratio Lower Upper
83 100
85 55.4 51.1 76.7
127 5.9 7.2 10.8#



#48
Methyl methacrylate
Concen: 2.276 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Tgt Ion: 41 Resp: 5712
Ion Ratio Lower Upper
41 100
69 83.7 56.3 84.5
39 62.9 50.3 75.5





#49

1,4-Dioxane

Concen: 48.347 ug/l

RT: 9.706 min Scan# 11

Delta R.T. 0.012 min


Lab File: VN083260.D

Acq: 13 Aug 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion: 88 Resp: 2189

Ion Ratio Lower Upper

88 100

43 30.5 27.8 41.8

58 71.3 59.4 89.0



#50

Toluene-d8

Concen: 50.640 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. 0.006 min

Lab File: VN083260.D




Acq: 13 Aug 2024 16:07


Tgt Ion: 98 Resp: 338557

Ion Ratio Lower Upper

98 100


100 65.1 51.5 77.3








#51
4-Methyl-2-Pentanone
Concen: 12.648 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 43 Resp: 36307
Ion Ratio Lower Upper
43 100
58 37.5 28.2 42.4



#52
Toluene
Concen: 2.582 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07


Tgt Ion: 92 Resp: 13175
Ion Ratio Lower Upper
92 100
91 172.4 139.4 209.0








#53
 t-1,3-Dichloropropene
 Concen: 2.356 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. -0.000 min
 Lab File: VN083260.D
 Acq: 13 Aug 2024 16:07

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion: 75 Resp: 7457
 Ion Ratio Lower Upper
 75 100
 77 29.2 25.7 38.5



#54
 cis-1,3-Dichloropropene
 Concen: 2.271 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. -0.000 min
 Lab File: VN083260.D
 Acq: 13 Aug 2024 16:07


Tgt Ion: 75 Resp: 7645
 Ion Ratio Lower Upper
 75 100
 77 39.2 24.3 36.5#
 39 58.8 50.5 75.7








#55
 1,1,2-Trichloroethane
 Concen: 2.683 ug/l
 RT: 11.018 min Scan# 1541
 Delta R.T. 0.000 min
 Lab File: VN083260.D
 Acq: 13 Aug 2024 16:07

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion:	97	Resp:	4906
Ion	Ratio	Lower	Upper
97	100		
83	72.7	71.4	107.0
85	55.0	45.8	68.6
99	61.2	52.0	78.0



#56
 Ethyl methacrylate
 Concen: 2.287 ug/l
 RT: 10.876 min Scan# 1517
 Delta R.T. 0.000 min
 Lab File: VN083260.D
 Acq: 13 Aug 2024 16:07

Tgt Ion:	69	Resp:	7886
Ion	Ratio	Lower	Upper
69	100		
41	71.8	63.4	95.2
39	49.5	37.4	56.0





#57

1,3-Dichloropropane

Concen: 2.559 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.000 min

Lab File: VN083260.D

Acq: 13 Aug 2024 16:07

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 76 Resp: 8343

Ion Ratio Lower Upper

76 100

78 34.1 25.6 38.4



#58

2-Chloroethyl Vinyl ether

Concen: 12.458 ug/l

RT: 10.165 min Scan# 1396

Delta R.T. 0.006 min

Lab File: VN083260.D




Acq: 13 Aug 2024 16:07


Tgt Ion: 63 Resp: 18155

Ion Ratio Lower Upper

63 100


106 27.9 21.4 32.0








#59
2-Hexanone
Concen: 11.910 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 43 Resp: 26450
Ion Ratio Lower Upper
43 100
58 54.7 24.6 73.8



#60
Dibromochloromethane
Concen: 2.452 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Tgt Ion: 129 Resp: 5422
Ion Ratio Lower Upper
129 100
127 74.3 39.2 117.6









#61
1,2-Dibromoethane
Concen: 2.508 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:107 Resp: 4819
Ion Ratio Lower Upper
107 100
109 87.9 73.7 110.5



#62
4-Bromofluorobenzene
Concen: 48.291 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Tgt Ion: 95 Resp: 125867
Ion Ratio Lower Upper
95 100
174 69.8 0.0 159.2
176 67.1 0.0 147.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:117	Resp:	245212
Ion	Ratio	Lower Upper
117	100	
82	61.6	47.5 71.3
119	32.1	25.6 38.4



#64
Tetrachloroethene
Concen: 2.597 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Tgt Ion:164	Resp:	4218
Ion	Ratio	Lower Upper
164	100	
166	126.1	101.4 152.0
129	80.0	75.7 113.5
131	97.2	73.8 110.6





#65

Chlorobenzene

Concen: 2.670 ug/l

RT: 11.894 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN083260.D

Acq: 13 Aug 2024 16:07

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:112 Resp: 14468

Ion Ratio Lower Upper

112 100

114 30.5 26.2 39.2



#66

1,1,1,2-Tetrachloroethane

Concen: 2.435 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.006 min

Lab File: VN083260.D

Acq: 13 Aug 2024 16:07




Tgt Ion:131 Resp: 4653


Ion Ratio Lower Upper

131 100

133 104.5 47.3 142.0

119 71.3 32.5 97.4









#67
Ethyl Benzene
Concen: 2.536 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 25209
Ion Ratio Lower Upper
91 100
106 27.3 24.0 36.0



#68
m/p-Xylenes
Concen: 4.902 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Tgt Ion: 106 Resp: 18254
Ion Ratio Lower Upper
106 100
91 213.9 166.1 249.1









#69
o-Xylene
Concen: 2.350 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:106 Resp: 8632
Ion Ratio Lower Upper
106 100
91 232.9 111.3 333.9



#70
Styrene
Concen: 2.392 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Tgt Ion:104 Resp: 14755
Ion Ratio Lower Upper
104 100
78 53.3 41.6 62.4
103 54.6 44.0 66.0





#71

Bromoform

Concen: 2.539 ug/l

RT: 12.582 min Scan# 11

Delta R.T. 0.006 min


Lab File: VN083260.D

Acq: 13 Aug 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion:173 Resp: 3676

Ion Ratio Lower Upper

173 100

175 46.0 24.0 72.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. -0.000 min

Lab File: VN083260.D

Acq: 13 Aug 2024 16:07




Tgt Ion:152 Resp: 109492


Ion Ratio Lower Upper

152 100

115 63.3 30.6 91.6

150 157.4 0.0 348.6









#73
Isopropylbenzene
Concen: 2.466 ug/l
RT: 12.700 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:105 Resp: 22585
Ion Ratio Lower Upper
105 100
120 25.3 12.9 38.6



#74
N-ethyl acetate
Concen: 2.250 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Tgt Ion: 43 Resp: 10079
Ion Ratio Lower Upper
43 100
70 43.0 29.8 44.6
55 24.3 18.7 28.1
61 22.5 17.4 26.0









#75
1,1,2,2-Tetrachloroethane
Concen: 2.638 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. 0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 6834
Ion Ratio Lower Upper
83 100
131 8.5 5.0 14.9
85 60.8 31.9 95.7



#76
1,2,3-Trichloropropane
Concen: 3.753 ug/l
RT: 12.994 min Scan# 1877
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Tgt Ion: 75 Resp: 9018
Ion Ratio Lower Upper
75 100
77 143.3 110.9 332.6





#77

Bromobenzene

Concen: 2.678 ug/l

RT: 12.976 min Scan# 1874

Delta R.T. -0.006 min

Lab File: VN083260.D





Acq: 13 Aug 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion:	156	Resp:	5448
Ion Ratio	Lower	Upper	
156	100		
77	237.2	111.3	333.8
158	96.0	48.3	144.9



#78

n-propylbenzene

Concen: 2.518 ug/l




RT: 13.035 min Scan# 1884


Delta R.T. 0.000 min

Lab File: VN083260.D

Acq: 13 Aug 2024 16:07


Tgt Ion:	91	Resp:	26554
Ion Ratio	Lower	Upper	
91	100		
120	21.5	10.9	32.9








#79
2-Chlorotoluene
Concen: 2.566 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion: 91 Resp: 17153
Ion Ratio Lower Upper
91 100
126 32.2 16.2 48.6




#80
1,3,5-Trimethylbenzene
Concen: 2.463 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07




Tgt Ion: 105 Resp: 18888
Ion Ratio Lower Upper
105 100
120 48.3 24.3 72.8








#81
trans-1,4-Dichloro-2-butene
Concen: 2.113 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion	Ratio	Lower	Upper
75	100		
53	104.6	88.4	132.6
89	37.1	36.3	54.5




#82
4-Chlorotoluene
Concen: 2.468 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07




Tgt Ion	Ratio	Lower	Upper
91	100		
126	31.8	16.2	48.6








#83
tert-Butylbenzene
Concen: 2.411 ug/l
RT: 13.435 min Scan# 1953
Delta R.T. -0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:119 Resp: 16371
Ion Ratio Lower Upper
119 100
91 73.5 35.0 105.1
134 29.3 13.3 39.8



#84
1,2,4-Trimethylbenzene
Concen: 2.384 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07


Tgt Ion:105 Resp: 18418
Ion Ratio Lower Upper
105 100
120 44.7 21.9 65.8








#85
 sec-Butylbenzene
 Concen: 2.395 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN083260.D
 Acq: 13 Aug 2024 16:07

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion:105 Resp: 22194
 Ion Ratio Lower Upper
 105 100
 134 19.5 9.8 29.3



#86
 p-Isopropyltoluene
 Concen: 2.239 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN083260.D
 Acq: 13 Aug 2024 16:07

Tgt Ion:119 Resp: 17129
 Ion Ratio Lower Upper
 119 100
 134 27.8 13.0 39.0
 91 27.7 12.3 36.9









#87
1,3-Dichlorobenzene
Concen: 2.572 ug/l
RT: 13.735 min Scan# 2003
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:146 Resp: 9844
Ion Ratio Lower Upper
146 100
111 42.9 20.6 61.9
148 60.6 31.8 95.3



#88
1,4-Dichlorobenzene
Concen: 2.551 ug/l
RT: 13.735 min Scan# 2003
Delta R.T. -0.076 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Tgt Ion:146 Resp: 9844
Ion Ratio Lower Upper
146 100
111 42.9 20.1 60.3
148 60.6 32.2 96.6





#89

n-Butylbenzene

Concen: 2.211 ug/l

RT: 14.059 min Scan# 2058

Delta R.T. -0.000 min

Lab File: VN083260.D





Acq: 13 Aug 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion:	91	Resp:	14656
Ion	Ratio	Lower	Upper
91	100		
92	53.0	26.1	78.3
134	24.4	12.1	36.3



#90

Hexachloroethane

Concen: 2.512 ug/l




RT: 14.329 min Scan# 2104


Delta R.T. -0.006 min

Lab File: VN083260.D

Acq: 13 Aug 2024 16:07


Tgt Ion:	117	Resp:	3712
Ion	Ratio	Lower	Upper
117	100		
201	51.4	35.8	107.3








#91
1,2-Dichlorobenzene
Concen: 2.458 ug/l
RT: 14.112 min Scan# 2067
Delta R.T. 0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:146 Resp: 9107
Ion Ratio Lower Upper
146 100
111 45.6 21.4 64.2
148 63.9 32.1 96.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 2.092 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. 0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07


Tgt Ion: 75 Resp: 1315
Ion Ratio Lower Upper
75 100
155 81.1 36.6 109.8
157 94.7 46.9 140.6








#93
1,2,4-Trichlorobenzene
Concen: 2.175 ug/l
RT: 15.394 min Scan# 21
Delta R.T. -0.000 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion:180 Resp: 4513
Ion Ratio Lower Upper
180 100
182 97.4 47.3 141.8
145 34.1 15.7 47.1




#94
Hexachlorobutadiene
Concen: 2.151 ug/l
RT: 15.494 min Scan# 2302
Delta R.T. -0.006 min
Lab File: VN083260.D
Acq: 13 Aug 2024 16:07



Tgt Ion:225 Resp: 1985
Ion Ratio Lower Upper
225 100
223 64.8 31.9 95.7
227 65.5 32.5 97.5





#95

Naphthalene

Concen: 2.052 ug/l

RT: 15.647 min Scan# 2327

Delta R.T. 0.006 min

Lab File: VN083260.D





Acq: 13 Aug 2024 16:07

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion:128	Resp:	15084
Ion Ratio	Lower	Upper
128	100	
127	12.5	10.6 16.0
129	10.1	8.6 12.8



#96

1,2,3-Trichlorobenzene

Concen: 2.145 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.000 min

Lab File: VN083260.D

Acq: 13 Aug 2024 16:07

Tgt Ion:180	Resp:	4405
Ion Ratio	Lower	Upper
180	100	
182	100.0	48.9 146.8
145	34.4	16.8 50.4

