

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
 Data File : VN083256.D
 Acq On : 13 Aug 2024 14:31
 Operator : JC\MD
 Sample : P3429-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-073124

Quant Time: Aug 14 01:44:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	147065	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	284471	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	291862	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	130851	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	116829	55.811	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 111.620%		
35) Dibromofluoromethane	8.171	113	92438	52.060	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.120%		
50) Toluene-d8	10.565	98	352167	53.170	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 106.340%		
62) 4-Bromofluorobenzene	12.847	95	148227	57.403	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 114.800%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
Data File : VN083256.D
Acq On : 13 Aug 2024 14:31
Operator : JC\MD
Sample : P3429-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-073124

Quant Time: Aug 14 01:44:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

