

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
 Data File : VN083262.D
 Acq On : 13 Aug 2024 16:56
 Operator : JC\MD
 Sample : P3440-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-080124

Quant Time: Aug 14 01:45:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	142679	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	276180	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	283985	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	127173	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	114991	56.621	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.240%
35) Dibromofluoromethane	8.165	113	91146	52.873	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.740%
50) Toluene-d8	10.565	98	342790	53.308	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.620%
62) 4-Bromofluorobenzene	12.847	95	143057	57.064	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	114.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
Data File : VN083262.D
Acq On : 13 Aug 2024 16:56
Operator : JC\MD
Sample : P3440-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-080124

Quant Time: Aug 14 01:45:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

