

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
 Data File : VN083264.D
 Acq On : 13 Aug 2024 17:44
 Operator : JC\MD
 Sample : P3548-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RBR200036

Quant Time: Aug 14 01:46:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	137896	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	270941	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	282618	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	127700	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	109360	55.717	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.440%
35) Dibromofluoromethane	8.171	113	87977	52.022	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.040%
50) Toluene-d8	10.565	98	335091	53.118	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.240%
62) 4-Bromofluorobenzene	12.853	95	139669	56.790	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	113.580%
Target Compounds						
						Qvalue
16) Acetone	4.441	43	24698	32.157	ug/l	98
20) Methylene Chloride	5.288	84	6309	3.569	ug/l	87
43) Isopropyl Acetate	8.688	43	137967	26.583	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
Data File : VN083264.D
Acq On : 13 Aug 2024 17:44
Operator : JC\MD
Sample : P3548-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RBR200036

Quant Time: Aug 14 01:46:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

