

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
 Data File : VN083265.D
 Acq On : 13 Aug 2024 18:08
 Operator : JC\MD
 Sample : P3548-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 343

Quant Time: Aug 14 01:47:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	141314	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	272555	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	284304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	128981	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	113557	56.455	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 112.920%		
35) Dibromofluoromethane	8.171	113	88964	52.294	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.580%		
50) Toluene-d8	10.565	98	340415	53.642	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 107.280%		
62) 4-Bromofluorobenzene	12.847	95	143637	58.058	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 116.120%		
Target Compounds						
						Qvalue
16) Acetone	4.436	43	24064	30.574	ug/l	90
20) Methylene Chloride	5.283	84	5691	3.142	ug/l #	72
43) Isopropyl Acetate	8.688	43	143902	27.643	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
Data File : VN083265.D
Acq On : 13 Aug 2024 18:08
Operator : JC\MD
Sample : P3548-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
343

Quant Time: Aug 14 01:47:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

