

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
 Data File : VN083279.D
 Acq On : 14 Aug 2024 05:40
 Operator : JC\MD
 Sample : P3450-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 922-K1-WP0-0.25-080124

Quant Time: Aug 14 07:58:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	138109	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	273792	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	284536	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	128029	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	111247	56.590	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.180%
35) Dibromofluoromethane	8.171	113	90185	52.772	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.540%
50) Toluene-d8	10.565	98	340598	53.429	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.860%
62) 4-Bromofluorobenzene	12.847	95	142707	57.421	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	114.840%
Target Compounds						
						Qvalue
29) Tetrahydrofuran	7.853	42	2438	3.193	ug/l	# 61
52) Toluene	10.635	92	5960	1.225	ug/l	98
86) p-Isopropyltoluene	13.729	119	4739	0.530	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
Data File : VN083279.D
Acq On : 14 Aug 2024 05:40
Operator : JC\MD
Sample : P3450-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
922-K1-WP0-0.25-080124

Quant Time: Aug 14 07:58:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

