

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
 Data File : VN083281.D
 Acq On : 14 Aug 2024 06:28
 Operator : JC\MD
 Sample : P3467-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 919-J-WS-080224

Quant Time: Aug 14 07:59:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	133782	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	261677	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	268472	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	116504	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	108932	57.205	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.420%
35) Dibromofluoromethane	8.165	113	86716	53.091	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.180%
50) Toluene-d8	10.565	98	329361	54.058	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.120%
62) 4-Bromofluorobenzene	12.847	95	133342	56.137	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	112.280%
Target Compounds						Qvalue
13) Acrolein	4.189	56	10570	40.982	ug/l	97
16) Acetone	4.448	43	10236	13.737	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
Data File : VN083281.D
Acq On : 14 Aug 2024 06:28
Operator : JC\MD
Sample : P3467-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
919-J-WS-080224

Quant Time: Aug 14 07:59:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

