

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
 Data File : VN083283.D
 Acq On : 14 Aug 2024 07:16
 Operator : JC\MD
 Sample : P3451-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 921-J-WS-080124

Quant Time: Aug 14 07:59:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	130924	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	257809	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	269897	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	123166	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	106234	57.006	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 114.020%	
35) Dibromofluoromethane	8.165	113	84689	52.628	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.260%	
50) Toluene-d8	10.565	98	319644	53.250	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.500%	
62) 4-Bromofluorobenzene	12.847	95	135397	57.857	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 115.720%	
Target Compounds						
						Qvalue
13) Acrolein	4.189	56	1551	6.145	ug/l	84
16) Acetone	4.436	43	7802	10.699	ug/l	93
52) Toluene	10.635	92	9096	1.985	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
Data File : VN083283.D
Acq On : 14 Aug 2024 07:16
Operator : JC\MD
Sample : P3451-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
921-J-WS-080124

Quant Time: Aug 14 07:59:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

