

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
 Data File : VN083284.D
 Acq On : 14 Aug 2024 07:40
 Operator : JC\MD
 Sample : P3451-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-03-080124

Quant Time: Aug 14 08:14:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	138901	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	271745	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	276868	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	120328	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	112204	56.752	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 113.500%	
35) Dibromofluoromethane	8.165	113	88634	52.255	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.520%	
50) Toluene-d8	10.565	98	337136	53.284	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.560%	
62) 4-Bromofluorobenzene	12.847	95	137725	55.834	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 111.660%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
Data File : VN083284.D
Acq On : 14 Aug 2024 07:40
Operator : JC\MD
Sample : P3451-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-03-080124

Quant Time: Aug 14 08:14:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

