

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
 Data File : VN083285.D
 Acq On : 14 Aug 2024 08:04
 Operator : JC\MD
 Sample : P3534-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PIPE-BLAST-GRIT(FB)

Quant Time: Aug 14 08:20:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	131144	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	257886	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	264947	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	116832	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	107634	57.660	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.320%
35) Dibromofluoromethane	8.171	113	84650	52.588	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.180%
50) Toluene-d8	10.565	98	313963	52.288	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.580%
62) 4-Bromofluorobenzene	12.847	95	130777	55.867	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	111.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
Data File : VN083285.D
Acq On : 14 Aug 2024 08:04
Operator : JC\MD
Sample : P3534-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PIPE-BLAST-GRIT(FB)

Quant Time: Aug 14 08:20:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

