

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
 Data File : VN083271.D
 Acq On : 13 Aug 2024 20:34
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 14 01:50:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	80	0.00
2 T	Dichlorodifluoromethane	0.567	0.536	5.5	76	0.00
3 P	Chloromethane	0.581	0.539	7.2	73	0.00
4 C	Vinyl Chloride	0.592	0.596	-0.7#	80	0.00
5 T	Bromomethane	0.368	0.273	25.8#	62	-0.01
6 T	Chloroethane	0.371	0.385	-3.8	87	0.00
7 T	Trichlorofluoromethane	0.979	1.023	-4.5	83	0.00
8 T	Diethyl Ether	0.364	0.365	-0.3	80	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.539	0.539	0.0	79	0.00
10 T	Methyl Iodide	0.710	0.554	22.0	60	0.00
11 T	Tert butyl alcohol	0.148	0.162	-9.5	89	0.00
12 CM	1,1-Dichloroethene	0.554	0.525	5.2#	77	0.00
13 T	Acrolein	0.096	0.092	4.2	72	0.00
14 T	Allyl chloride	1.047	0.878	16.1	76	0.00
15 T	Acrylonitrile	0.304	0.312	-2.6	82	0.00
16 T	Acetone	0.278	0.263	5.4	74	0.00
17 T	Carbon Disulfide	1.622	1.364	15.9	70	0.00
18 T	Methyl Acetate	0.830	0.882	-6.3	91	0.00
19 T	Methyl tert-butyl Ether	2.001	2.039	-1.9	81	0.00
20 T	Methylene Chloride	0.641	0.610	4.8	82	0.00
21 T	trans-1,2-Dichloroethene	0.573	0.559	2.4	78	-0.01
22 T	Diisopropyl ether	1.969	2.037	-3.5	82	0.00
23 T	Vinyl Acetate	2.018	2.125	-5.3	81	0.00
24 P	1,1-Dichloroethane	1.073	1.113	-3.7	83	0.00
25 T	2-Butanone	0.428	0.428	0.0	81	0.00
26 T	2,2-Dichloropropane	0.997	0.812	18.6	64	0.00
27 T	cis-1,2-Dichloroethene	0.691	0.699	-1.2	82	0.00
28 T	Bromochloromethane	0.439	0.450	-2.5	82	0.00
29 T	Tetrahydrofuran	0.276	0.287	-4.0	83	0.00
30 C	Chloroform	1.115	1.201	-7.7#	85	0.00
31 T	Cyclohexane	1.055	0.946	10.3	77	0.00
32 T	1,1,1-Trichloroethane	1.055	1.107	-4.9	83	0.00
33 S	1,2-Dichloroethane-d4	0.712	0.849	-19.2	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.312	0.346	-10.9	87	0.00
36 T	1,1-Dichloropropene	0.472	0.460	2.5	81	0.00
37 T	Ethyl Acetate	0.529	0.493	6.8	83	0.00
38 T	Carbon Tetrachloride	0.532	0.533	-0.2	82	0.00
39 T	Methylcyclohexane	0.580	0.539	7.1	76	0.00
40 TM	Benzene	1.406	1.397	0.6	81	0.00
41 T	Methacrylonitrile	0.301	0.278	7.6	82	0.00
42 TM	1,2-Dichloroethane	0.512	0.537	-4.9	86	0.00
43 T	Isopropyl Acetate	1.211	0.960	20.7	87	0.00
44 TM	Trichloroethene	0.335	0.328	2.1	80	0.00
45 C	1,2-Dichloropropane	0.334	0.346	-3.6#	84	0.00
46 T	Dibromomethane	0.239	0.249	-4.2	83	0.00
47 T	Bromodichloromethane	0.537	0.553	-3.0	86	0.00
48 T	Methyl methacrylate	0.437	0.414	5.3	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
 Data File : VN083271.D
 Acq On : 13 Aug 2024 20:34
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 14 01:50:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	80	0.00
50 S	Toluene-d8	1.164	1.302	-11.9	86	0.00
51 T	4-Methyl-2-Pentanone	0.500	0.517	-3.4	84	0.00
52 CM	Toluene	0.889	0.901	-1.3#	81	0.00
53 T	t-1,3-Dichloropropene	0.551	0.551	0.0	77	0.00
54 T	cis-1,3-Dichloropropene	0.586	0.568	3.1	78	0.00
55 T	1,1,2-Trichloroethane	0.318	0.334	-5.0	84	0.00
56 T	Ethyl methacrylate	0.600	0.607	-1.2	81	0.00
57 T	1,3-Dichloropropane	0.568	0.584	-2.8	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.254	0.240	5.5	77	0.00
59 T	2-Hexanone	0.387	0.393	-1.6	83	0.00
60 T	Dibromochloromethane	0.385	0.411	-6.8	83	0.00
61 T	1,2-Dibromoethane	0.335	0.345	-3.0	83	0.00
62 S	4-Bromofluorobenzene	0.454	0.514	-13.2	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	85	0.00
64 T	Tetrachloroethene	0.331	0.320	3.3	79	0.00
65 PM	Chlorobenzene	1.105	1.086	1.7	81	0.00
66 T	1,1,1,2-Tetrachloroethane	0.390	0.392	-0.5	83	0.00
67 C	Ethyl Benzene	2.027	2.000	1.3#	82	0.00
68 T	m/p-Xylenes	0.759	0.755	0.5	81	0.00
69 T	o-Xylene	0.749	0.742	0.9	81	0.00
70 T	Styrene	1.258	1.281	-1.8	81	0.00
71 P	Bromoform	0.295	0.304	-3.1	81	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	4.182	3.921	6.2	82	0.00
74 T	N-amyl acetate	2.046	1.832	10.5	81	0.00
75 P	1,1,2,2-Tetrachloroethane	1.183	1.153	2.5	88	0.00
76 T	1,2,3-Trichloropropane	1.097	1.025	6.6	86	0.00
77 T	Bromobenzene	0.929	0.902	2.9	84	0.00
78 T	n-propylbenzene	4.816	4.590	4.7	82	0.00
79 T	2-Chlorotoluene	3.053	2.872	5.9	83	0.00
80 T	1,3,5-Trimethylbenzene	3.501	3.353	4.2	83	0.00
81 T	trans-1,4-Dichloro-2-butene	0.504	0.403	20.0	73	0.00
82 T	4-Chlorotoluene	3.062	2.912	4.9	84	0.00
83 T	tert-Butylbenzene	3.101	2.909	6.2	82	0.00
84 T	1,2,4-Trimethylbenzene	3.529	3.385	4.1	83	0.00
85 T	sec-Butylbenzene	4.231	3.983	5.9	81	0.00
86 T	p-Isopropyltoluene	3.493	3.396	2.8	82	0.00
87 T	1,3-Dichlorobenzene	1.748	1.677	4.1	84	0.00
88 T	1,4-Dichlorobenzene	1.762	1.659	5.8	83	0.00
89 T	n-Butylbenzene	3.027	2.921	3.5	81	0.00
90 T	Hexachloroethane	0.675	0.631	6.5	82	0.00
91 T	1,2-Dichlorobenzene	1.692	1.632	3.5	85	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.287	0.261	9.1	83	0.00
93 T	1,2,4-Trichlorobenzene	0.948	0.885	6.6	80	0.00
94 T	Hexachlorobutadiene	0.421	0.358	15.0	78	0.00
95 T	Naphthalene	3.357	3.093	7.9	79	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081324\
Data File : VN083271.D
Acq On : 13 Aug 2024 20:34
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 14 01:50:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.938	0.855	8.8	79	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6