

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081723\
 Data File : VN078925.D
 Acq On : 17 Aug 2023 14:39
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

Reviewed By :John Carlone 08/18/2023
 Supervised By :Semsettin Yesilyurt 08/18/2023

Quant Time: Aug 18 05:24:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N081723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Aug 18 05:21:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.824	128	43116m	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	234932	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	197842	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	103504	29.631	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	98.767%
60) 4-Bromofluorobenzene	12.853	95	71095	24.635	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	82.100%
63) Toluene-d8	10.571	98	283161	30.648	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.167%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081723\
Data File : VN078925.D
Acq On : 17 Aug 2023 14:39
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

Reviewed By : John Carlone 08/18/2023
Supervised By : Semsettin Yesilyurt 08/18/2023

Quant Time: Aug 18 05:24:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N081723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Aug 18 05:21:14 2023
Response via : Initial Calibration

