



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081723\
Data File : VN078917.D
Acq On : 17 Aug 2023 09:20
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N081723W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Fri Aug 18 05:21:14 2023



AutoFind: Scans 1852, 1853, 1854; Background Corrected with Scan 1845

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.0	9598	PASS
75	95	30	60	58.2	24280	PASS
95	95	100	100	100.0	41744	PASS
96	95	5	9	8.9	3702	PASS
173	174	0.00	2	0.5	154	PASS
174	95	50	100	80.2	33459	PASS
175	174	5	9	7.5	2511	PASS
176	174	95	101	96.8	32405	PASS
177	176	5	9	7.6	2449	PASS