

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
 Data File : VN083368.D
 Acq On : 19 Aug 2024 12:21
 Operator : JC\MD
 Sample : VN0819MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0819MBL01

Quant Time: Aug 20 04:43:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	111668	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	220576	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	230647	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	103931	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	97792	61.525	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 123.040%	
35) Dibromofluoromethane	8.165	113	80454	58.436	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 116.880%	
50) Toluene-d8	10.565	98	290917	56.645	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 113.300%#	
62) 4-Bromofluorobenzene	12.847	95	117344	58.607	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 117.220%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
Data File : VN083368.D
Acq On : 19 Aug 2024 12:21
Operator : JC\MD
Sample : VN0819MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0819MBL01

Quant Time: Aug 20 04:43:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

