

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN081924\  
 Data File : VN083379.D  
 Acq On : 19 Aug 2024 17:53  
 Operator : JC\MD  
 Sample : P3657-01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 917-J-WS-081624

Quant Time: Aug 20 04:49:26 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N080724W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Aug 08 06:30:41 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min)  |
|-----------------------------|--------|-------|----------|----------|-------|-----------|
| -----                       |        |       |          |          |       |           |
| Internal Standards          |        |       |          |          |       |           |
| 1) Pentafluorobenzene       | 8.224  | 168   | 139398   | 50.000   | ug/l  | 0.00      |
| 34) 1,4-Difluorobenzene     | 9.106  | 114   | 273365   | 50.000   | ug/l  | 0.00      |
| 63) Chlorobenzene-d5        | 11.865 | 117   | 283963   | 50.000   | ug/l  | 0.00      |
| 72) 1,4-Dichlorobenzene-d4  | 13.794 | 152   | 126838   | 50.000   | ug/l  | 0.00      |
| System Monitoring Compounds |        |       |          |          |       |           |
| 33) 1,2-Dichloroethane-d4   | 8.582  | 65    | 105845   | 53.345   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =     | 106.680%  |
| 35) Dibromofluoromethane    | 8.171  | 113   | 83464    | 48.916   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 97.840%   |
| 50) Toluene-d8              | 10.565 | 98    | 311490   | 48.939   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =     | 97.880%   |
| 62) 4-Bromofluorobenzene    | 12.853 | 95    | 127359   | 51.326   | ug/l  | 0.00      |
| Spiked Amount               | 50.000 | Range | 77 - 121 | Recovery | =     | 102.660%  |
| Target Compounds            |        |       |          |          |       |           |
| 16) Acetone                 | 4.447  | 43    | 5705     | 7.348    | ug/l  | Qvalue 95 |
| -----                       |        |       |          |          |       |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN081924\  
Data File : VN083379.D  
Acq On : 19 Aug 2024 17:53  
Operator : JC\MD  
Sample : P3657-01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
917-J-WS-081624

Quant Time: Aug 20 04:49:26 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N080724W.M  
Quant Title : SW846 8260  
QLast Update : Thu Aug 08 06:30:41 2024  
Response via : Initial Calibration

