

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
 Data File : VN083381.D
 Acq On : 19 Aug 2024 18:42
 Operator : JC\MD
 Sample : P3645-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 916-J-WS-081524

Quant Time: Aug 20 04:50:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	132595	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	260707	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	265791	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	115723	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	95901	50.813	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.620%
35) Dibromofluoromethane	8.171	113	79014	48.556	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.120%
50) Toluene-d8	10.570	98	289541	47.699	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.400%
62) 4-Bromofluorobenzene	12.847	95	118225	49.958	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
Data File : VN083381.D
Acq On : 19 Aug 2024 18:42
Operator : JC\MD
Sample : P3645-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
916-J-WS-081524

Quant Time: Aug 20 04:50:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

