

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
 Data File : VN083383.D
 Acq On : 19 Aug 2024 19:30
 Operator : JC\MD
 Sample : P3646-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 916-J-WPO-0.25-081524

Quant Time: Aug 20 04:50:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	136272	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	268732	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	278824	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	125925	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	103472	53.345	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.680%			
35) Dibromofluoromethane	8.171	113	83030	49.500	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.000%			
50) Toluene-d8	10.570	98	304710	48.699	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 97.400%			
62) 4-Bromofluorobenzene	12.853	95	127222	52.154	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 104.300%			
Target Compounds						Qvalue
16) Acetone	4.447	43	2426	3.196	ug/l	94
29) Tetrahydrofuran	7.847	42	3034	4.027	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
Data File : VN083383.D
Acq On : 19 Aug 2024 19:30
Operator : JC\MD
Sample : P3646-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
916-J-WPO-0.25-081524

Quant Time: Aug 20 04:50:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

