

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
 Data File : VN083384.D
 Acq On : 19 Aug 2024 19:55
 Operator : JC\MD
 Sample : P3662-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 917-J-WPO-0.25-081624

Quant Time: Aug 20 04:51:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	134810	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	268344	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	281709	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	122958	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	105905	55.191	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 110.380%		
35) Dibromofluoromethane	8.165	113	85294	50.923	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.840%		
50) Toluene-d8	10.571	98	312305	49.985	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.980%		
62) 4-Bromofluorobenzene	12.853	95	129453	53.146	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 106.300%		
Target Compounds						
						Qvalue
16) Acetone	4.436	43	12297	16.377	ug/l	# 86
25) 2-Butanone	7.494	43	14433	12.519	ug/l	98
29) Tetrahydrofuran	7.841	42	66771	89.592	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081924\
Data File : VN083384.D
Acq On : 19 Aug 2024 19:55
Operator : JC\MD
Sample : P3662-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
917-J-WPO-0.25-081624

Quant Time: Aug 20 04:51:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

