

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082123\
 Data File : VN078967.D
 Acq On : 21 Aug 2023 11:55
 Operator : JC\MD
 Sample : VN0821MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0821MBL01

Quant Time: Aug 22 07:40:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	160467	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	287156	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	256000	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	76240	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	118573	47.486	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.980%
35) Dibromofluoromethane	8.171	113	100445	51.964	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.920%
50) Toluene-d8	10.571	98	355578	49.542	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.080%
62) 4-Bromofluorobenzene	12.853	95	98194	41.856	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082123\
Data File : VN078967.D
Acq On : 21 Aug 2023 11:55
Operator : JC\MD
Sample : VN0821MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0821MBL01

Quant Time: Aug 22 07:40:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

