

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082123\
 Data File : VN078976.D
 Acq On : 21 Aug 2023 15:44
 Operator : JC\MD
 Sample : 04082-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 217

Quant Time: Aug 22 00:58:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	189105	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	352327	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	345953	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	108004	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	147171	50.013	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.020%	
35) Dibromofluoromethane	8.171	113	122288	51.562	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.120%	
50) Toluene-d8	10.571	98	443353	50.345	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.700%	
62) 4-Bromofluorobenzene	12.853	95	135003	46.902	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.800%	
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	39365	32.421	ug/l #	82
20) Methylene Chloride	5.277	84	10467	4.045	ug/l #	78
25) 2-Butanone	7.506	43	10937	5.618	ug/l #	66
43) Isopropyl Acetate	8.700	43	184175	26.105	ug/l #	75
51) 4-Methyl-2-Pentanone	10.453	43	41973	10.031	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082123\
Data File : VN078976.D
Acq On : 21 Aug 2023 15:44
Operator : JC\MD
Sample : 04082-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
217

Quant Time: Aug 22 00:58:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

