

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082124\
 Data File : VN083420.D
 Acq On : 21 Aug 2024 19:55
 Operator : JC\MD
 Sample : P3669-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20430

Quant Time: Aug 22 08:12:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	139352	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	290694	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	314860	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	154935	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	108613	54.758	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.520%
35) Dibromofluoromethane	8.165	113	86790	47.833	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.660%
50) Toluene-d8	10.565	98	319006	47.132	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.260%
62) 4-Bromofluorobenzene	12.847	95	147666	55.962	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	111.920%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	13634	17.566	ug/l	98
18) Methyl Acetate	5.030	43	9440	4.082	ug/l	97
20) Methylene Chloride	5.265	84	7859	4.400	ug/l #	78
43) Isopropyl Acetate	8.688	43	212991	39.322	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082124\
Data File : VN083420.D
Acq On : 21 Aug 2024 19:55
Operator : JC\MD
Sample : P3669-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20430

Quant Time: Aug 22 08:12:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

