

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082222\
 Data File : VN074044.D
 Acq On : 22 Aug 2022 13:47
 Operator : JC\MD
 Sample : VN0822WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0822WBL01

Quant Time: Aug 22 14:08:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	482130	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	739794	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	712325	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	409891	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	239770	51.059	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.120%	
35) Dibromofluoromethane	7.951	113	233543	53.381	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.760%	
50) Toluene-d8	10.380	98	779461	46.001	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 92.000%	
62) 4-Bromofluorobenzene	12.668	95	339252	60.328	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 120.660%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082222\
Data File : VN074044.D
Acq On : 22 Aug 2022 13:47
Operator : JC\MD
Sample : VN0822WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0822WBL01

Quant Time: Aug 22 14:08:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

