

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082222\
 Data File : VN074048.D
 Acq On : 22 Aug 2022 16:42
 Operator : JC\MD
 Sample : N4247-01RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-1RE

Quant Time: Aug 22 17:11:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	489636	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	750157	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	748323	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	424209	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	247803	51.961	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.920%	
35) Dibromofluoromethane	7.951	113	238925	53.856	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.720%	
50) Toluene-d8	10.380	98	824728	48.001	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.000%	
62) 4-Bromofluorobenzene	12.669	95	357811	62.749	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 125.500%#	
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	68490	30.440	ug/l	99
20) Methylene Chloride	5.016	84	36756	7.742	ug/l	96
25) 2-Butanone	7.269	43	17859	5.006	ug/l	96
37) Ethyl Acetate	7.334	43	71841	9.624	ug/l	97
40) Benzene	8.392	78	57903	2.977	ug/l	99
43) Isopropyl Acetate	8.492	43	284022	26.707	ug/l #	92
52) Toluene	10.445	92	58444	4.382	ug/l	100
68) m/p-Xylenes	11.892	106	30444	2.833	ug/l	100
69) o-Xylene	12.216	106	20672	1.918	ug/l	97
84) 1,2,4-Trimethylbenzene	13.316	105	40400	1.610	ug/l	65
95) Naphthalene	15.433	128	439764	14.802	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082222\
Data File : VN074048.D
Acq On : 22 Aug 2022 16:42
Operator : JC\MD
Sample : N4247-01RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-1RE

Quant Time: Aug 22 17:11:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

