

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082323\
 Data File : VN079018.D
 Acq On : 23 Aug 2023 16:00
 Operator : JC\MD
 Sample : 04091-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-H

Quant Time: Aug 24 01:15:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	172165	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	334797	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	335264	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	99612	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	146320	54.617	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.240%	
35) Dibromofluoromethane	8.171	113	117550	52.159	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.320%	
50) Toluene-d8	10.570	98	427999	51.146	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.300%	
62) 4-Bromofluorobenzene	12.853	95	132865	48.576	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.160%	
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	27053	24.473	ug/l #	79
20) Methylene Chloride	5.283	84	13992	5.940	ug/l #	73
43) Isopropyl Acetate	8.694	43	199392	29.741	ug/l #	71
51) 4-Methyl-2-Pentanone	10.447	43	36502	9.180	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082323\
Data File : VN079018.D
Acq On : 23 Aug 2023 16:00
Operator : JC\MD
Sample : 04091-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-H

Quant Time: Aug 24 01:15:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

