

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082324\
 Data File : VN083454.D
 Acq On : 23 Aug 2024 12:35
 Operator : JC\MD
 Sample : PB162920TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162920TB

Quant Time: Aug 23 23:58:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	131737	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	267287	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	300055	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	140114	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	103673	55.289	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	110.580%	
35) Dibromofluoromethane	8.165	113	80820	48.443	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	96.880%	
50) Toluene-d8	10.565	98	303938	48.838	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	97.680%	
62) 4-Bromofluorobenzene	12.847	95	138775	57.198	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	114.400%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	10314	14.057	ug/l	91
20) Methylene Chloride	5.271	84	3372	1.997	ug/l #	90
43) Isopropyl Acetate	8.688	43	167589	33.267	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082324\
Data File : VN083454.D
Acq On : 23 Aug 2024 12:35
Operator : JC\MD
Sample : PB162920TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162920TB

Quant Time: Aug 23 23:58:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

