

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082324\
 Data File : VN083455.D
 Acq On : 23 Aug 2024 12:59
 Operator : JC\MD
 Sample : PB162920ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162920ZHE#01

Quant Time: Aug 24 01:33:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	127214	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	264439	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	289384	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	134179	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	106341	58.728	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	117.460%	
35) Dibromofluoromethane	8.165	113	83682	50.699	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.400%	
50) Toluene-d8	10.565	98	313627	50.938	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	101.880%	
62) 4-Bromofluorobenzene	12.847	95	140416	58.498	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	117.000%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	10479	14.790	ug/l #	89
18) Methyl Acetate	5.030	43	5346	2.532	ug/l	100
20) Methylene Chloride	5.259	84	3700	2.269	ug/l #	83
43) Isopropyl Acetate	8.688	43	217787	44.573	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082324\
Data File : VN083455.D
Acq On : 23 Aug 2024 12:59
Operator : JC\MD
Sample : PB162920ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162920ZHE#01

Quant Time: Aug 24 01:33:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

