

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082324\
 Data File : VN083459.D
 Acq On : 23 Aug 2024 14:36
 Operator : JC\MD
 Sample : P3694-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Kingsland-Driveway

Quant Time: Aug 24 00:01:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	131423	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	271817	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	297008	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	141633	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	106560	56.964	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.920%
35) Dibromofluoromethane	8.165	113	85230	50.235	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.480%
50) Toluene-d8	10.565	98	315923	49.918	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.840%
62) 4-Bromofluorobenzene	12.847	95	142883	57.910	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	115.820%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	32083	43.830	ug/l	99
18) Methyl Acetate	5.030	43	6608	3.030	ug/l	97
20) Methylene Chloride	5.265	84	5290	3.140	ug/l #	66
43) Isopropyl Acetate	8.688	43	138357	26.571	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082324\
Data File : VN083459.D
Acq On : 23 Aug 2024 14:36
Operator : JC\MD
Sample : P3694-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Kingsland-Driveway

Quant Time: Aug 24 00:01:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

