

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082324\
 Data File : VN083461.D
 Acq On : 23 Aug 2024 15:31
 Operator : JC\MD
 Sample : P3715-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK-240821052-04

Quant Time: Aug 24 00:02:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	119726	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	247116	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	259622	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	108622	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	97327	57.111	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.220%
35) Dibromofluoromethane	8.165	113	75777	49.128	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.260%
50) Toluene-d8	10.565	98	284148	49.385	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.780%
62) 4-Bromofluorobenzene	12.847	95	116686	52.020	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.040%
Target Compounds						
18) Methyl Acetate	5.024	43	3020	1.520	ug/l	Qvalue # 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082324\
Data File : VN083461.D
Acq On : 23 Aug 2024 15:31
Operator : JC\MD
Sample : P3715-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK-240821052-04

Quant Time: Aug 24 00:02:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

