

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082324\
 Data File : VN083462.D
 Acq On : 23 Aug 2024 15:55
 Operator : JC\MD
 Sample : P3716-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1048-MW-01(23)

Quant Time: Aug 24 00:03:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	122781	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	249387	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	276626	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	139204	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	100945	57.760	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 115.520%		
35) Dibromofluoromethane	8.165	113	79482	51.061	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.120%		
50) Toluene-d8	10.565	98	283527	48.829	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 97.660%		
62) 4-Bromofluorobenzene	12.853	95	140304	61.979	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 123.960%#		
Target Compounds						
						Qvalue
4) Vinyl Chloride	2.512	62	30346	20.863	ug/l	98
16) Acetone	4.447	43	4322	6.320	ug/l #	72
18) Methyl Acetate	5.024	43	4969	2.439	ug/l #	78
27) cis-1,2-Dichloroethene	7.482	96	247981	146.087	ug/l	84
44) Trichloroethene	9.353	130	211377	126.597	ug/l	92
64) Tetrachloroethene	11.106	164	1669432	911.249	ug/l	93
67) Ethyl Benzene	11.970	91	3909	0.349	ug/l	96
68) m/p-Xylenes	12.070	106	6058	1.442	ug/l	95
69) o-Xylene	12.400	106	1693	0.409	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082324\
Data File : VN083462.D
Acq On : 23 Aug 2024 15:55
Operator : JC\MD
Sample : P3716-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1048-MW-01(23)

Quant Time: Aug 24 00:03:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

