

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090123\
 Data File : VN079121.D
 Acq On : 01 Sep 2023 14:34
 Operator : JC\MD
 Sample : PB155232TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB155232TB

Quant Time: Sep 01 23:12:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	258530	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	477576	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	449158	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	133529	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	185270	46.053	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.100%
35) Dibromofluoromethane	8.171	113	165757	51.561	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.120%
50) Toluene-d8	10.571	98	588840	49.330	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.660%
62) 4-Bromofluorobenzene	12.853	95	185902	47.647	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.300%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	28424	17.124	ug/l	90
20) Methylene Chloride	5.289	84	25115	7.100	ug/l #	91
43) Isopropyl Acetate	8.694	43	181137	18.941	ug/l #	84
51) 4-Methyl-2-Pentanone	10.447	43	33755	5.951	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090123\
Data File : VN079121.D
Acq On : 01 Sep 2023 14:34
Operator : JC\MD
Sample : PB155232TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB155232TB

Quant Time: Sep 01 23:12:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

