

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090223\
 Data File : VN079142.D
 Acq On : 02 Sep 2023 14:47
 Operator : JC\MD
 Sample : PB155301ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB155301ZHE#04

Quant Time: Sep 04 01:04:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	204590	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	398355	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	387909	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	114183	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	154857	48.642	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.280%
35) Dibromofluoromethane	8.177	113	135640	50.583	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.160%
50) Toluene-d8	10.571	98	514089	51.632	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.260%
62) 4-Bromofluorobenzene	12.853	95	153456	47.153	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.300%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	28367	21.595	ug/l #	85
20) Methylene Chloride	5.283	84	29752	10.628	ug/l	92
37) Ethyl Acetate	7.577	43	4669	1.035	ug/l #	83
43) Isopropyl Acetate	8.694	43	196693	24.658	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090223\
Data File : VN079142.D
Acq On : 02 Sep 2023 14:47
Operator : JC\MD
Sample : PB155301ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB155301ZHE#04

Quant Time: Sep 04 01:04:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

