

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090223\
 Data File : VN079149.D
 Acq On : 02 Sep 2023 17:35
 Operator : JC\MD
 Sample : 04256-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2

Quant Time: Sep 04 01:07:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	265694	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	506062	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	480796	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	160644	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	201291	48.687	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.380%
35) Dibromofluoromethane	8.171	113	179067	52.566	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.140%
50) Toluene-d8	10.570	98	625109	49.420	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.840%
62) 4-Bromofluorobenzene	12.853	95	203736	49.278	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.560%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	30123	17.658	ug/l	94
20) Methylene Chloride	5.283	84	20279	5.578	ug/l #	79
39) Methylcyclohexane	9.606	83	5062	1.074	ug/l	92
43) Isopropyl Acetate	8.694	43	263934	26.045	ug/l #	72
51) 4-Methyl-2-Pentanone	10.453	43	31635	5.263	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090223\
Data File : VN079149.D
Acq On : 02 Sep 2023 17:35
Operator : JC\MD
Sample : 04256-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2

Quant Time: Sep 04 01:07:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

