

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090223\
 Data File : VN079155.D
 Acq On : 02 Sep 2023 19:59
 Operator : JC\MD
 Sample : 04285-01 200X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 830

Quant Time: Sep 04 01:10:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	256239	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	481843	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	352592	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	143619	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	192108	48.180	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.360%
35) Dibromofluoromethane	8.171	113	168062	51.815	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.620%
50) Toluene-d8	10.571	98	571362	47.442	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.880%
62) 4-Bromofluorobenzene	12.853	95	186085	47.271	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.540%
Target Compounds						
					Qvalue	
20) Methylene Chloride	5.277	84	5598	1.597	ug/l	89
95) Naphthalene	15.647	128	7587	3.662	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090223\
Data File : VN079155.D
Acq On : 02 Sep 2023 19:59
Operator : JC\MD
Sample : 04285-01 200X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
830

Quant Time: Sep 04 01:10:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

