

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
 Data File : VN087754.D
 Acq On : 08 Sep 2025 12:30
 Operator : JC\MD
 Sample : Q3036-03 4000X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EGR-2

Quant Time: Sep 09 02:26:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	246375	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	570287	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	530434	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	251773	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	268989	53.287	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.580%
35) Dibromofluoromethane	8.147	113	205837	49.991	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.980%
50) Toluene-d8	10.547	98	745941	48.403	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.800%
62) 4-Bromofluorobenzene	12.829	95	254363	46.411	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.820%
Target Compounds						
20) Methylene Chloride	5.271	84	59	Below Cal	Qvalue #	53

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 09 02:26:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration

