

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
 Data File : VN087757.D
 Acq On : 08 Sep 2025 13:40
 Operator : JC\MD
 Sample : Q3023-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Sep 09 02:28:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	177578	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	409804	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	404049	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	188140	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	191072	52.516	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 105.040%		
35) Dibromofluoromethane	8.153	113	143729	48.577	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.160%		
50) Toluene-d8	10.541	98	544162	49.138	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 98.280%		
62) 4-Bromofluorobenzene	12.824	95	189781	48.188	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 96.380%		
Target Compounds						
						Qvalue
16) Acetone	4.418	43	47894	24.181	ug/l	100
18) Methyl Acetate	5.024	43	30668	7.395	ug/l	99
20) Methylene Chloride	5.265	84	10440	2.464	ug/l	89
25) 2-Butanone	7.471	43	26308	10.096	ug/l	99
37) Ethyl Acetate	7.553	43	19439	3.134	ug/l	99
43) Isopropyl Acetate	8.677	43	14204	1.467	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090825\
Data File : VN087757.D
Acq On : 08 Sep 2025 13:40
Operator : JC\MD
Sample : Q3023-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Sep 09 02:28:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration

