

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091125\
 Data File : VN087805.D
 Acq On : 11 Sep 2025 17:51
 Operator : JC\MD
 Sample : Q3083-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW4I

Quant Time: Sep 12 02:57:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	220322	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	503440	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	493455	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	228190	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	242746	53.775	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.540%	
35) Dibromofluoromethane	8.147	113	189798	52.217	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.440%	
50) Toluene-d8	10.547	98	653379	48.027	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.060%	
62) 4-Bromofluorobenzene	12.829	95	228728	47.276	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 94.560%	
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	80974	32.952	ug/l	99
22) Diisopropyl ether	6.665	45	4214	0.339	ug/l #	85
30) Chloroform	7.941	83	5171	0.744	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091125\
Data File : VN087805.D
Acq On : 11 Sep 2025 17:51
Operator : JC\MD
Sample : Q3083-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW4I

Quant Time: Sep 12 02:57:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration

