

Data Path : W:\HPCHEM1\MSVOA N\DATA\VN091515\
 Data File : VN027174.D
 Acq On : 15 Sep 2015 19:10
 Operator : MD\FY
 Sample : I.BLK
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 I.BLK

Quant Time: Sep 15 21:01:31 2015
 Quant Method : W:\HPCHEM1\MSVOA_N\METHODS\82N091515W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 15 17:32:41 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.75	168	825536	50.00	ug/l	0.00
35) 1,4-Difluorobenzene	8.68	114	1484304	50.00	ug/l	0.00
64) Chlorobenzene-d5	11.52	117	1234539	50.00	ug/l	0.00
73) 1,4-Dichlorobenzene-d4	13.47	152	432363	50.00	ug/l	0.00

System Monitoring Compounds						
34) 1,2-Dichloroethane-d4	8.12	65	825355	76.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	152.60%	
36) Dibromofluoromethane	7.68	113	540042	59.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	118.68%	
51) Toluene-d8	10.19	98	2154645	59.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	118.58%	
63) 4-Bromofluorobenzene	12.52	95	731245	61.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	122.90%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA N\DATA\VN091515\
Data File : VN027174.D
Acq On : 15 Sep 2015 19:10
Operator : MD\FY
Sample : I.BLK
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
I.BLK

Quant Time: Sep 15 21:01:31 2015
Quant Method : W:\HPCHEM1\MSVOA_N\METHODS\82N091515W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 15 17:32:41 2015
Response via : Initial Calibration

