

Data Path : W:\HPCHEM1\MSVOA N\DATA\VN091515\
 Data File : VN027186.D
 Acq On : 16 Sep 2015 2:34
 Operator : MD\FY
 Sample : G3629-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S39-0100

Quant Time: Sep 16 04:35:56 2015
 Quant Method : W:\HPCHEM1\MSVOA_N\METHODS\82N091515W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 15 21:12:09 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.75	168	790266	50.00	ug/l	0.00
35) 1,4-Difluorobenzene	8.68	114	1417028	50.00	ug/l	0.00
64) Chlorobenzene-d5	11.52	117	1173981	50.00	ug/l	0.00
73) 1,4-Dichlorobenzene-d4	13.47	152	406493	50.00	ug/l	0.00
System Monitoring Compounds						
34) 1,2-Dichloroethane-d4	8.12	65	778615	48.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.18%	
36) Dibromofluoromethane	7.68	113	509365	51.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.72%	
51) Toluene-d8	10.19	98	2053188	53.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.78%	
63) 4-Bromofluorobenzene	12.52	95	683564	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
Target Compounds						
4) Chloromethane	2.01	50	3966	1.48	ug/l #	79
5) Vinyl Chloride	2.14	62	164502	14.73	ug/l	100
20) Methyl tert-butyl Ether	5.13	73	8275	0.20	ug/l #	87
25) 1,1-Dichloroethane	5.94	63	33951	1.36	ug/l #	93
28) cis-1,2-Dichloroethene	6.92	96	132972	10.56	ug/l	91
66) Chlorobenzene	11.54	112	9742	0.35	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA N\DATA\VN091515\
Data File : VN027186.D
Acq On : 16 Sep 2015 2:34
Operator : MD\FY
Sample : G3629-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S39-0100

Quant Time: Sep 16 04:35:56 2015
Quant Method : W:\HPCHEM1\MSVOA_N\METHODS\82N091515W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 15 21:12:09 2015
Response via : Initial Calibration

