

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091522\
 Data File : VN074461.D
 Acq On : 15 Sep 2022 14:47
 Operator : JC\MD
 Sample : VN0915WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0915WBL01

Quant Time: Sep 15 19:22:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	494660	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	941593	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	891392	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	377690	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	354537	42.242	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.480%
35) Dibromofluoromethane	7.951	113	291698	46.636	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.280%
50) Toluene-d8	10.381	98	1075726	44.470	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.940%
62) 4-Bromofluorobenzene	12.675	95	414938	50.084	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091522\
Data File : VN074461.D
Acq On : 15 Sep 2022 14:47
Operator : JC\MD
Sample : VN0915WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0915WBL01

Quant Time: Sep 15 19:22:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

